The hand surface electrode HOF 2 consists of two point electrodes and a holding body of

The point-shaped electrodes are made of conductive silicone elastomer. By means of two integrated ports, the hand surface electrode HOF 2 was to be connected to the Milli-TO 3 and TO 3. The simple 4-mm sockets even allow connection to instruments from

The use of silicone elastomer, the surface resistance measurement of concave and

insulating material.

other manufacturers.

convex specimens is possible.



Measuring Electrode

for the measurement of surface resistance electrically conductive layers in the aerospace industry based on DIN 65181 according to AIRBUS Radom Test NTX-CMM 53-15-11



HOF 2 - point-shaped electrodes

- Surface resistance measurement according based on DIN 65181 according to AIRBUS Radom Test NTX-CMM 53-15-11
- optimized for measurements on concave and convex specimens in air and space
- two point electrodes made of conductive Silicone elastomer
- Electrode diameter of 4 mm
- Electrode distance 50 mm
- ▶ Following Milli-TO 3 and TO 3
- Resistance range 1 Ω bis 100 MΩ
- Measuring voltage to 500 volts





HOF 2 connection with Milli-TO 3

HOF 2 - 4mm connection sockets

H.-P. FISCHER ELEKTRONIK GmbH & Co. Industrie- und Labortechnik KG Karl-Metten-Ring 1 D-15749 Mittenwalde / GERMANY phone: fax: email: web: +49 (0)33764 25560 +49 (0)33764 255625 info@fischer-messtechnik.de www.fischer-messtechnik.de