

Exhalation Valve Leakage Test Equipment

Description

The exhalation valve leakage test equipment is as described in AS/NZS 1716. The equipment applies the specified pressure drop across the exhalation valve, and indicates the flow rate of the air through the valve in its closed position.



Services required

Bench mounted
Compressed air (maximum 2 bar)
24mm spanner
One minute timer

Approximate size & weight

50 x 50 x 50 cm: 2 kg

Relevant standard

AS/NZS 1716:2012 Appendix F