DIN EN 71-2

EN 71-2



SUPPLIES

Propane gas / natural gas Eleltrical current 230 VAC

TO BE PROVIDED BY THE CUSTOMER

Flue gas exhaust system DN200

Specimen holder 4.2.5



SCOPE

The test device according to EN 71-2 is used to determine the flammability of toys under specified test conditions.

PRINCIPLE

Specimens of various types are mounted on suitable specimen holders and flamed with a burner. A protocol documenting the characteristics of the fire test must be drawn up after all standard tests.

SPECIAL FEATURES

Electric fan


Electronic stopwatch with software

SCOPE OF DELIVERY

Test chamber with electrical fan and ventilation flap Burner with linear guide and mechanical fine control valve with numeric display Specimen holder EN 71-2 4.2.5, angle adjustable Specimen holder fixture with linear guide Specimen holder EN 71-1 4.2.2, Headdress Specimen holder EN 71-2 4.5, Soft fillings Drip tray Electronic stopwatch with software Flame gauge

DIMENSIONS

Width x depth x height: approx. 1186 x 893 x 1953 mm* Weight: approx. 100 kg





Specimen holder 4.2.2



* Our products are constantly evolving. For this reason, the actual d<mark>imensions may differ.</mark> © 01/2020



理寶科技有限公司 Libero Techonology Company Limited

T: (852) 2555 8222

香港 Hong Kong 上海 Shanghai

T: 86 (21) 5655 8285 廣州 Guangzhou T: 86 (20) 3928 3292

Email: sales@liberohk.com www.liberohk.com